

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, FOR THE MONTH OF AS ON-12.07.16.

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipments : 3524 (Updated up to 30.06.16)
 3. Total Number of Non-Functional Equipments : 16
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	One	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	In ref. to the letter dt.11.05.16 of Dy. Director (PC), comments of HOD(Radiology) has been received and same has been forwarded to Dy. Director (PC) on 11.07.16 for further needful.
2	Radiotherapy	HDR Brachytherapy machine	One			December,2009	No	Technical restrictions from AERB		Letter dated-09.7.16 sent to R.S.O. with approval of M.S. for some clarification.
3	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	-----	Repair order placed to the firm on 17.06.16.
4	Blood Bank	Deep Freezer M/s. Yorco Scientific	01	March 2014	6 Lac	Dec.2015	Under warranty	Compressor not working	-----	Firm representative has assured either to repair or replace the material shortly & extend the warranty period accordingly.
5	Blood Bank	Deep Freezer M/s. Yorco Scientific	01	March 2014	6 Lac	April, 16	No	Compressor not working	-----	Firm representative has assured either to repair or replace the material shortly & extend the warranty period accordingly.
6	Medicine	Ventilator M/s. Instromedix	02	August,2009	4.00 lac each	January,2016	No	Spares required	Spares required	File still lying with HOD (Medicine) and is under discussion with

										departmental committee.
7	Dental	Sand Blaster M/s. IDS Denmed	01	23.02.10	2.50 lacs	2011	No	Firm has not provided the service during warranty period	Spares required	Repair estimate submitted by the firm. Meeting for negotiation of price with the firm is fixed for 15.07.16.
8	Dental	Ceramic Furnace M/s. International Dental System	01	1.07.10	5.04 lacs	2011	No	- do -	- do -	- do -
9	Dental	Duplicating Unit M/s. IDS Denmed	01	01.07.10	2.01 lacs	2011	No	- do -	- do -	- do -
10	Dental	Apex Locator M/s. IDS Denmed	02	23.02.10	1.28 lacs	2011	No	- do -	- do -	Firm has submitted the inspection report and file has been sent to M.S. with comments of HOD for further direction.
11	Pediatrics Ward-18	Transport Ventilator Elisee-150 M/s. Instromedix	01	2009	3.95 lacs	December,2015	No.	Spares required	Turbine, Rotary valve	Spares have been received and under installation.
12	Pediatrics Ward-20	Transport Ventilator M/s. Instromedix	01	2009	3.95 lac	January,16	No	Spares required	Turbine, Rotary valve, Peep motor, oxygen sensor	Repair order placed on to the firm on 13.06.2016.
13	Microbiology	CO2 Incubator M/s. Yorco sales	01	2014	4.93 lac	April,16	Yes	-----	-----	Letter sent to firm on 10.06.16 and reminder on 27.06.16 to replace the machine. Reply received from the firm and same has been put up to HOD (Micro) for perusal and comments.
14	Biochemistry	Automatic Clinical Chemistry Analyzer	01	19.01.10	17.58 lacs	-----	No	Cooling unit not working	Estimate for cooling unit Rs.10.55 lac.	Repair estimate found non-economical. Process initiated for condemnation.

sd/
C.M.O. (Medical Stores)