

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTINS MORE THAN RS. ONE LAKH, FOR THE MONTH OF –May, 2014

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
2. Total Number of Equipment : 1569
3. Total Number of Non-Functional equipment : 02
4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioningof equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1.	Pediatrics Ward-18	Bear Cub Ventilator M/s. Rohanika Electronics	One	2009	USD 14520	March,14	No	Spares required	Flow Sensor	Functional
2	Eye	Scrub Station M/s. Rudex International	Two	2010	1,90,000/-	April, 14	No	Spares required	Sencer, foot pump, etc.	Letter dt.19.05.14 & Reminder-I dt.06.06.14 sent to firm to clarify the queries raised by the repair committee.

Sd/-
C.M.O. (Medical Store)