

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RS. ONE LAKH, FOR THE MONTH OF MARCH-2015

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipment : 3061 (updated up to 31.03.15)
 3. Total Number of Non-Functional equipment : 04
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non-functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Radiology	1000 MA Digital X-Ray Machine M/s .Empire Instrumentation	1	2007	Euro 296000	Frequently gone out order and now continuously non functional.	Under Warranty	Some parts to be changed	Spares required	Letter received from the principal firm M/s. Italray informing that since M/s. Empire closed its medical equipment business hence they have no control over them and if desired italray can submit a quotation for visit of their engineer to S.J.Hospital. Letter sent to Director (PC) Dt.GHS on-18.04.15 for necessary action.
2.	Anesthesia	Elisee 150 Ventilator M/s. Instromedix	1	2009	3.95 lac	Sent to firm's workshop for repair on 26.8.14, but not returned back, till date.	No	Calibration required	Calibration required	Repair estimate of Rs.3.50 lacs submitted by the firm was put up to the repair committee, but not approved by the committee as the estimate is non-economical in comparison to the cost of the equipment. File will be sent to HOD for comments.

3.	S.I.C.	Xenon Light Source M/s. Vishal Surgical Equipiment co.	2	20.01.11	8961 Euro	April,2015	AMC under process	Light cable not working	Fiber optic light cable	Estimate submitted by the firm, file sent to HOD on 20.4.15 for verification of estimate.
----	--------	--	---	----------	-----------	------------	----------------------	----------------------------	----------------------------	--

C.M.O. (Medical Store)