

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, FOR THE MONTH OF AS ON-14.06.16.

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipments : 3441 (Updated up to 31.03.16)
 3. Total Number of Non-Functional Equipments : 15
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	One	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	Letter dated-11.05.16 received from Dy. Director (Proc.). File has been sent to HOD (Radiology) for comments. File not received yet. Reminder sent to HOD on 14.06.16.
2	Radiotherapy	HDR Brachytherapy machine	One			December,2009	No	Technical restrictions from AERB		Company representative was asked to clarify rates for CMC which have not been quoted earlier. CMC quotation received and verified by HOD and meeting has been called between HOD and R.S.O. to finalize the matter.
3	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	-----	Estimate approved by the Repair committee. File has been put up to the M.S. for Expenditure sanction.
4	Resp. Medicine	Video Thorow scope M/s. Olympus Medical	01	26.06.12	33.00 lacs	Nov, 2015	Yes	Spares required	-----	Functional
5	Blood Bank	Deep Freezer M/s. Yorco Scientific	01	March 2014	6 Lac	Dec.2015	Under warranty	Compressor not working	-----	Firm has intimated that they have placed the order for spare parts to the principal firm and will repair the machine shortly. Reminder sent to the firm on 8.04.16. Letter again sent to the firm for show cause notice. Another

										letter received from the firm saying spares are lying in the custom and equipment will be repaired soon.
6	Blood Bank	Deep Freezer M/s. Yorco Scientific	01	March 2014	6 Lac	April, 16	No	Compressor not working	-----	Complaint received from HOD. Letter has been sent to the firm for repair.
7	Medicine	Ventilator M/s. Instromedix	02	August,2009	4.00 lac each	January,2016	No	Spares required	Spares required	File again sent to HOD (Med.) last week, with required documents & original purchase file. Reminder sent to HOD on 14..06.16 to expedite the matter.
8	Dental	Sand Blaster M/s. IDS Denmed	01	23.02.10	2.50 lacs	2011	No	Firm has not provided the service during warranty period	Spares required	Firm has submitted the estimate and file has been sent to HOD (Dental) for filling up the requisite forms.
9	Dental	Ceramic Furnace M/s. International Dental System	01	1.07.10	5.04 lacs	2011	No	- do -	- do -	- do -
10	Dental	Duplicating Unit M/s. IDS Denmed	01	01.07.10	2.01 lacs	2011	No	- do -	- do -	- do -
11	Dental	Apex Locator M/s. IDS Denmed	02	23.02.10	1.28 lacs	2011	No	- do -	- do -	Firm has submitted the inspection report and file has been sent to HOD for his comments.
12	Pediatrics Ward-18	Transport Ventilator Elisee-150 M/s. Instromedix	01	2009	3.95 lacs	December,2015	No.	Spares required	Turbine, Rotary valve	Repair order placed to the firm on 09.05.16. Reminder has been sent to the firm.
13	Pediatrics Ward-20	Transport Ventilator M/s. Instromedix	01	2009	3.95 lac	January,16	No	Spares required	Turbine, Rotary valve, Peep motor, oxygen sensor	Repair order placed on to the firm on 13.06.2016.
14	Microbiology	CO2 Incubator M/s. Yorco sales	01	2014	4.93 lac	April,16	Yes	-----	-----	Letter sent to firm on 10.06.16 to replace the equipment.
15	Microbiology	Vertical Autoclave M/s. Yorco Sales	01	2013	20.00 lac	April,16	Yes	-----	-----	Functional

Sd/-
C.M.O. (Medical Store)

