

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, FOR THE MONTH OF AS ON-09.08.16.

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipments : 3524 (Updated up to 30.06.16)
 3. Total Number of Non-Functional Equipments : 15
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	One	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	In ref. to the letter dt.11.05.16 of Dy. Director (PC), comments of HOD(Radiology) has been received and same has been forwarded to Dy. Director (PC) on 11.07.16 for further needful.
2	Radiotherapy	HDR Brachytherapy machine	One			December,2009	No	Technical restrictions from AERB		Letter dated-09.7.16 sent to R.S.O. with approval of M.S. for some clarification. File still with HOD for R.S.O. signature. Reminder given on- 06.8.16
3	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	-----	Repair order placed to the firm on 17.06.16. Reminder to the firm on-08.08.16.
4	Blood Bank	Deep Freezer M/s. Yorco Scientific	01	March 2014	6 Lac	Dec.2015	Under warranty	Compressor not working	-----	Competent authority has decided to conduct a meeting under the chairmanship of Addl.M.S., HOD and firm representative on-22.08.16 to decide the matter.
5	Blood Bank	Deep Freezer M/s. Yorco Scientific	01	March 2014	6 Lac	April, 16	No	Compressor not working	-----	- do -

6	Medicine	Ventilator M/s. Instrumedix	02	August,2009	4.00 lac each	January,2016	No	Spares required	Spares required	File received back from HOD with remarks that repair cost is on higher side. Letter to be sent to the firm for negotiation.
7	Dental	Sand Blaster M/s. IDS Denmed	01	23.02.10	2.50 lacs	2011	No	Firm has not provided the service during warranty period	Spares required	Meeting for negotiation of price was held on 15.07.16. Firm representative was asked to submit revised quotations with lower prices, if possible with clause of warranty. The reply from firm is awaited. A reminder in this regard has been sent to firm on 30.07.16
8	Dental	Ceramic Furnace M/s. International Dental System	01	1.07.10	5.04 lacs	2011	No	- do -	- do -	- do -
9	Dental	Duplicating Unit M/s. IDS Denmed	01	01.07.10	2.01 lacs	2011	No	- do -	- do -	- do -
10	Dental	Apex Locator M/s. IDS Denmed	02	23.02.10	1.28 lacs	2011	No	- do -	- do -	Letter written to the firm on 27.07.16 for clarification regarding the equipment.
11	Pediatrics Ward-18	Transport Ventilator Elisee- 150 M/s. Instrumedix	01	2009	3.95 lacs	December,2015	No.	Spares required	Turbine, Rotary valve	Spares have been received and under installation. Firm has taken the equipment at their workshop last week for calibration.
12	Pediatrics Ward-20	Transport Ventilator M/s. Instrumedix	01	2009	3.95 lac	January,16	No	Spares required	Turbine, Rotary valve, Peep motor, oxygen sensor	Repair order placed on to the firm on 13.06.2016. Reminder sent on 08.08.16.
13	Microbiology	CO2 Incubator M/s. Yorco Scientific	01	2014	4.93 lac	April,16	Yes	-----	-----	Competent authority has decided to conduct a meeting under the chairmanship of Addl.M.S., HOD and firm representative on-22.08.16 to decide the matter.

Sd/-
C.M.O. (Medical Store)