

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RS. ONE LAKH, FOR THE MONTH OF FEBRUARY-2015

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipment : 2900 (updated up to 31.12.14)
 3. Total Number of Non-Functional equipment : 02
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non- functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	ENT	O.T.Table M/s. Hospimedica International	1	2001	15.50 lacs	November,14	AMC	Spares required	Spares required	Functional
2	Radiology	1000 MA Digital X- Ray Machine M/s .Empire Instrumentation	1	2007	Euro 296000	Frequently gone out order and now continuously non functional.	Under Warranty	Some parts to be changed	Spares required	Letter has been sent to the firm on 4.12.14 to make the machine functional, immediately. Reminder sent on 19.12.14 and 12.01.14, 04.02.15. Letter received from the firm informing that the division has been shut down. Letter sent to principal firm on 9.03.15
3.	Anesthesia	Elisee 150 Ventilator M/s. Instromedix	1	2009	3.95 lac	Sent to firm's workshop for repair on 26.8.14, but not returned back, till date.	No	Calibration required	Calibration required	Firm has returned back the ventilator unrepaired and submitted repair estimate, file has been sent to HOD for further needful.

SD/-
C.M.O. (Medical Store)