

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, FOR THE MONTH OF AS ON-11.07.17.

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipments : 3849 (Updated up to 31.03.17)
 3. Total Number of Non-Functional Equipments : 09
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	01	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	The user department has sent their recommendations to the chairman, condemnation committee for condemnation of the equipment on 29.05.2017. Next condemnation committee is supposed to be held in mid of July, 2017. A letter to Dte.G.H.S. for blacklisting of the foreign principal has been sent on 03.07.2017.
2	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	-----	Supply Order to the firm to repair the equipment issued on 01.06.2017 with the condition to provide the backup equipment. The time given to the firm to supply these spares is 12.07.2017. The firm has collected the defective equipment for repair at their service centre and provided a backup equipment on 12.06.2017. A letter has been written to the firm for an early repair of the equipment on 05.07.2017, but spares not

										received till date.
3	Medicine	Ventilator M/s. Instromedix	04	August,2009	4.00 lac each	January,2016	No	Spares required	Spares required	Inspection done on 15.06.2017, File sent to user department for installation on 23.06.2017, An office note has also sent to HOD Medicine for urgent actions for repair of the equipment on 07.07.2017, but no reply received till date.
4	Dental	Apex Locator M/s. IDS Denmed New firm M/s. Hosta care	02	23.02.10	1.28 lacs	2011	No	- do -	- do -	M/s. Hosta care has submitted the quotation for repair. The file has been sent to HOD Dental on 25.04.17 to examine the estimate. File is yet to be received back from HOD. Reminders have been sent to HOD on 17.05.17, 09.06.17 and 05.07.17, but no reply received till date.
5	Pathology	Fully Automated Blood cell counter M/s. CPC Diagnostic	01	17.08.2007	4.62 lacs	-----	Under warranty at the time of equipment become non functional	Not repaired by the firm	Not provided by the firm	The file has been sent to user department on 24.04.17 to initiate and follow up the condemnation process of the equipment and process of blacklisting of the firm. User department has informed that the equipment is under process for condemnation. A letter has been sent to the HOD to get the present status on 12.06.2017 and 06.07.2017 but no reply received till date.

C.M.O. (Medical Store)