

**STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, AS ON- 13.08.2019.**

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI  
 2. Total Number of Equipments : 3849 (Updated up to 31.03.17)  
 3. Total Number of Non-Functional Equipments : 03  
 4. Theft or Fire Occurred : Nil

| 1     | 2              | 3                                   | 4   | 5                                 | 6                          | 7  | 8  | 9  | 10  | 11   |
|-------|----------------|-------------------------------------|---|-----------------------------------|----------------------------|--|--|--|---|--|
| S.No. | Name of Deptt. | Name of equipment Department wise   | Total No. of equipment & date of purchase | Date of installation of equipment | Cost of equipment (Rs.)    | Date from which equipment(s) is/are non - functional | Whether the equipment(s) is under AMC/warranty, if not reasons thereof | Reasons for non functioning of equipment | Details of part required for repair along with estimate expenditure | Action taken to repair the equipment & approximate time by which equipment will become functional  |
| 1     | PMR            | CAD- CAM                            | 01<br>15.09.2008                          | July 2009                         | Rs.<br>62,00,000/-         | Feb., 2016   | Under CMC  | Loss of USB Key                          | USB Key<br><br>Approximate expenditure:<br>Rs. 4,78,000/-           | Omega CAD software successfully installed and working satisfactory. However, problem in the functioning of scanner camera and carver system. Firm has inspected the machine. Letter dt. 06.08.19 already issued to firm to submit the repair estimate. |
| 2     | PMR            | Ultrasound Therapy Model: Medison   | 01<br>01.10.2010                          | Jan, 2011                         | Rs. 1,17,600/-             | Functional   | No   | Functional                               | Functional  | Functional   |
| 3.    | Microbiology   | Horizontal Steam Sterilizer-YSU 608 | 01<br>25.10.2011                          | 03.01.2012                        | Rs.17,12,550/-             | October, 2018  | No   | Technical                                | -<br><br>Approximate expenditure:<br>Rs. 65,000/-                   | Query has been resolved by the firm and file has been submitted to Inspection committee on 09.08.19.   |
| 4.    | Eye            | USG AB Scan                         | -   | 18.07.2013                        | Rs. 4,20,000/-<br>+ 5% Tax | June, 2019   | No   | Defect in Probe Cable                    | -<br><br>Approximate expenditure:<br>Rs. 50,000/-                   | Supply received inspection done. File sent to HOD for installation report.   |

**C.M.O. (Medical Store)**