

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, March, 2018.

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipments : 3849 (Updated up to 31.03.17)
 3. Total Number of Non-Functional Equipments : NIL + 02(to be condemned)
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non - functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	-----	The firm has repaired the machine and returned the same to CTVS department on 20.02.2018
2	Pathology	Fully Automated Blood cell counter M/s. CPC Diagnostic	01	17.08.2007	4.62 lacs	-----	Under warranty at the time of equipment become non functional	Not repaired by the firm	Not provided by the firm	To be Condemned.
3	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	01	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	To be Condemned.

C.M.O. (Medical Store)