

STATUS OF NON-FUNCTIONAL EQUIPMENTS COSTING MORE THAN RUPEES ONE LAKH, FOR THE MONTH OF AS ON-16.05.17.

1. Name of Hospital : V.M.M.C. & SAFDARJANG HOSPITAL, NEW DELHI
 2. Total Number of Equipments : 3849 (Updated up to 31.03.17)
 3. Total Number of Non-Functional Equipments : 10
 4. Theft or Fire Occurred : Nil

1	2	3	4	5	6	7	8	9	10	11
S.No.	Name of Deptt.	Name of equipment Department wise	Total No. of equipment & date of purchase	Date of installation of equipment	Cost of equipment (Rs.)	Date from which equipment(s) is/are non-functional	Whether the equipment(s) is under AMC/warranty, if not reasons thereof	Reasons for non functioning of equipment	Details of part required for repair along with estimate expenditure	Action taken to repair the equipment & approximate time by which equipment will become functional
1	Radiology	1000MA Digital X-Ray Machine M/s. Empire Instrumentation	01	2007	Euro 296000	Frequently gone out of order and now continuously non functional	No	Some parts to be changed	Spares required	The file is lying with the user department (Radiology). It is telephonically informed by the user deptt. on 10.04.2017 that they have started the condemnation process, and will provide the files as and when their work with the file will finished. User department (Radiology) has been requested to return the file and provide the updated status of the condemnation and blacklisting of the firm vide office note dated 12.04.17 and 05.05.17, but no reply received till date.
2	CTVS	Cardiac Output Monitor M/s. Edwards Life Sciences	01	2009	6.12 lacs	Dec, 2015	No	Spares required	----	Firm has re-inspected the equipment and submitted a revised estimate of repair with request for permission to collect the equipment to send it to their Japan Tech. services. Repair Committee has approved the proposal. The firm has been requested vide letters Dt. 14.02.17, 03.03.17, 10.04.2017 and 08.05.17 (also emailed) to deposit collateral security towards

										cost of equipment. No reply has been received.
3	Medicine	Ventilator M/s. Instromedix	04	August,2009	4.00 lac each	January,2016	No	Spares required	Spares required	Spares for the repair of ventilators (04 Nos.) have been received in the store on 15.05.17 and repair is under process.
4	Dental	Apex Locator M/s. IDS Denmed New firm M/s. Hosta care	02	23.02.10	1.28 lacs	2011	No	- do -	- do -	M/s. Hosta care has submitted the quotation for repair. The file has been sent to HOD Dental on 25.04.17 to examine the estimate. File is yet to be received back.
5	Blood Bank	Automatic cell counter M/s. Beckman coulter	01	2014	10.39 lacs	Nov, 16	Under warranty till 19.12.16	-----	-----	Letter sent to firm on 07.12.16 for repair. Reminder sent to the firm on 03.01.17 and 02/17. A reminder has been sent to the firm on 04.05.17 for inspection and repair of the equipment on urgent basis.
6	Pathology	Fully Automated Blood cell counter M/s. CPC Diagnostic	01	17.08.2007	4.62 lacs	-----	Under warranty at the time of equipment become non functional	Not repaired by the firm	Not provided by the firm	An inspection was done by Dte.G.H.S. representative (Dr. Anil Sain, Addl .Dy. Dir. General) in April,2017 and advised the user department to initiate the process of blacklist the firm and look into the possibilities of getting the equipment condemned as the equipment is 10 years old and it is difficult to get the equipment repaired from the foreign firm. Accordingly the file has been sent to user department on 24.04.17 to take necessary action.

C.M.O. (Medical Store)